

Determination of the Geographic Origin of Spices Using Elemental Analysis after Microwave Digestion

Lindsey Whitecotton¹, Greg Gilleland¹, Jenny Nelson¹⁻³, Courtney Tanabe^{2,3}, Elaine Hasty⁴, Leanne Anderson⁴,
¹Agilent Technologies, Inc., 5301 Stevens Creek Blvd, Santa Clara, CA 95051
²Food Safety and Measurement Facility, University of California, Davis, One Shields Ave., Davis CA 95616
³Department of Viticulture and Enology, University of California, Davis, One Shields Ave., Davis, CA 95616
⁴CEM Corporation, 3100 Smith Farm Road, Matthews, NC 28104

Pittcon
210-46

Introduction

Competitive advantage within the food industry has lead to fraudulent activities like intentional counterfeiting, substitution, adulteration, or mislabeling/ misrepresentation of ingredients within products. Many of these tricks go unnoticed by consumers and regulating government agencies due in part to the lack of standardized methods for identification. This highly specialized problem results in inferior products entering the marketplace and skewing price points for legitimate goods. One good example of this is Madagascar vanilla, price per kilo has vaulted from \$47 in 2012/13 to \$385 in 2016/17. This leads to potential counterfeiting and substitution from bad actors/criminals in the food industry. For origin-driven products like spices, food fraud has highlighted the need for ways to defend claims of geographic origin. This study uses three different technologies; microwave digestion, ICP-OES, and ICP-MS, to determine if spice origin from different countries could be accurately distinguished. Spice samples were sourced from a large US importer of spices that can verify the country of origin. The results of this study demonstrate the validity and applicability of the analytical approaches used.



Experimental

Samples. Over 50 spices from around the world were used in this study. All acquired from a Business-to-Business spice company in the US.

Sample Preparation. All spice samples were digested in a MARS 6 with MARSXpress Plus Vessels.

Experimental



Figure 1: MARSXpress Plus vessels (left). Spices and the vessels (middle), group photo (right).

"Plant Material" One Touch Method MARSXpress Plus Vessels	
Stages	1
Power	290 - 1800
Ramp Time	20:00
Hold Time	10:00
Temperature	200 °C

Table 1: Microwave Program Specs. with the control points for the Plant Material program that we used for all digestions (left), Screen shot from a run showing all 24 vessels are at the same temperature, and that the batch reached and held the programmed 200 °C temperature for the required 10 minutes (right).

Digested samples were then prepared for elemental analysis:

- Samples were roughly 100x dilution (0.5g into 50mL).
- Final concentration was approximately 20% acid (samples were digested in 9mL HNO3 and 1mL HCl, for a final acid concentration 18%HNO3 and 2%HCl).

Instrument parameters. An Agilent 7900 ICP-MS with the Ultra High Matrix Introduction (UHMI) system was used. The sample introduction system consisting of a micromist concentric nebulizer, quartz spray chamber, quartz torch with 2.5 mm id injector, and platinum cones.

The 5110 ICP-OES sample introduction system consisted of a SeaSpray nebulizer, double-Pass cyclonic spray chamber, and a 1.8 mm id injector torch.

Instrument operating conditions are listed in Table 2.



Table 2. ICP-MS/ICP-OES instrument operating conditions

ICP-MS Parameter	Setting	ICP-OES Parameter	Setting
RF power (W)	1600	RF power (kW)	1.20 kW
Sampling depth (mm)	10	Auxiliary gas flow (L/min)	1.0 L/min
Carrier gas flow (L/min)	0.68	Plasma gas flow (L/min)	12.0 L/min
Dilution (HMI) gas flow (L/min)	0.27	Nebulizer gas flow (L/min)	0.7 L/min
Helium cell gas flow (mL/min)	4.3	Viewing mode	Axial/Radial
Hydrogen cell gas flow (mL/min)	5	Replicate Read Time (s)	30 sec/5 sec
Energy discrimination (V)	4	Viewing height (mm)	12 mm
		Switching Valve	AVS7

Reference Standards. Various SRMs purchased from National Institute of Standards and Technology (NIST) were analyzed in this study to verify the sample preparation digestion process. NIST 1573a Tomato Leaves, NIST 1515 Apple Leaves, and NIST 1575 Pine Needles.

Samples and Sample Preparation. 50+ spices were obtained from a spice company that does B2B sales. They sourced all of the spices, and therefore know the true country of origin. This is very important when doing authentication studies. The samples were digested following the same procedure as for NIST SRM samples. About 0.5 g of each sample was weighted into 50 mL microwave vessels.

Results and Discussion

To verify the digestion process of the spice samples, four SRMs were analyzed by ICP-MS and ICP-OES. The mean results shown in Table 3 were in good agreement with the certified concentrations, where provided.

ICP-MS & ICP-OES (n Italic)	NIST 1573a Tomato Leaves				NIST 1575 Pine Needles				NIST 1515 Apples Leaves			
Element (mass and wavelength)	Certified conc	Mean Measure d conc*	Recovery %**	QC Criteria (80- 120%) ***	Certified conc	Mean Measure d conc*	Recovery %**	QC Criteria (80- 120%) ***	Certified conc	Mean Measure d conc*	Recovery %**	QC Criteria (80- 120%) ***
²³ Na	136	111	82%	PASS	63*	59	94%					
²⁴ Mg (279.078 nm)	12000*	11390	95%	PASS	1060	887	84%	PASS				
²⁷ Al (396.152 nm)	598	476	80%	PASS	580	537	93%	PASS				
³⁹ K (769.897 nm)	27000	28718	106%	PASS	4170	4209	101%	PASS				
⁴⁰ Ca (315.867 nm)	32500	49844	99%	PASS	2500	2475	99%	PASS				
⁴¹ K	0.835	0.687	82%	PASS					0.254	0.266	105%	PASS
⁵² Cr (205.560 nm)	99	1.70	86%	PASS								
⁵⁵ Mn (257.610 nm)	246	240	98%	PASS	488*	504	103%					
⁵⁶ Fe (239.553 nm)	389	359	92%	PASS	46	47	102%	PASS	82.7	72.2	87%	PASS
⁵⁹ Co (238.892 nm)	0.57	0.52	91%	PASS	1061*	0.065	100%					
⁶⁰ Ni (216.555 nm)	1.59	1.44	91%	PASS	1.47*	1.49	101%		0.936	0.774	83%	PASS
⁶³ Cu (324.754 nm)	4.7	4.0	85%	PASS	2.8	3.0	107%	PASS	5.69	4.79	84%	PASS
⁶⁶ Zn (213.857 nm)	30.9	25.5	82%	PASS	30.9	36.7	119%	PASS	12.45	10.40	84%	PASS
⁷⁵ As (188.950 nm)	0.112	0.119	106%	PASS	0.039*	0.032	82%					
⁷⁶ Se (196.026 nm)	0.054	0.061	113%	PASS	0.099*	0.103	104%					
⁸¹ Rb									10.2	8.65	84%	PASS
⁸³ Sr									25.1	23.6	94%	PASS
⁸⁷ Rb									0.095	0.078	82%	PASS
⁹⁰ Mo (202.032 nm)	0.46*	0.39	85%									
⁹¹ Zr	0.017*	0.013	75%						0.0132	0.0116	88%	PASS
⁹² Mo (226.502 nm)	0.52	0.51	98%	PASS	0.233	0.227	97%	PASS	0.0432	0.0493	114%	PASS
⁹³ Y	0.034	0.037	108%	PASS	6	6	100%	PASS	48.8	47.4	97%	PASS
⁹⁴ Zr	0.034	0.037	108%	PASS	6	6	100%	PASS	48.8	47.4	97%	PASS
⁹⁵ Mo	0.12*	0.10	80%		0.167*	0.166	99%		0.47	0.44	94%	PASS
⁹⁶ Zr	0.05*	0.023	46%									

*n=3, replicate sample digestion, each in triplicate

† FDA Elemental Analysis Manual (Section 3.4 Special Calculations) 3.4 Equation 20

** QC Criteria FDA EAM 47 (80-120%) for NIST certified values

*** Non-Certified Reference Value

n=3, replicate sample digestion, each in triplicate.
** QC Criteria FDA EAM 4.7 (80-120%) for NIST certified values
*** Non-Certified Reference Value

Results and Discussion

Spike Recoveries. Spike recovery test was carried out to check the accuracy of the ICP-MS and ICP-OES methods for actual sample analysis. Four random spice samples were spiked with all elements were spiked at 20 and 60 ppb.

Recoveries were between 82% (for Ba) and 119% (for Cu). Quantitative results for the spice samples showed that the concentrations of Al, K, Ca, Mg, Na, Fe, P, S, Si, Zn and Mn, were relatively high in all spiked spice sample. Using the 7900 ICP-MS and 5110 ICP-OES direct analysis methods, excellent spike recoveries were achieved for most elements in the spiked samples. All recoveries were within ±20% as shown in Table 4. The spike results for the elements listed above were invalid as the spike levels were much too low (20 times lower) relative to the levels present in the unspiked samples, therefore NA.

Table 4. Quantitative and spike recovery results for 4 spice samples.

Element (mass and wavelength)	White Pepper Sample 58		White Pepper Sample 45		Paprika Sample 24		Basil Sample 50		ICP-MS		ICP-OES	
	Measured conc, n=3, ppb								Mean Recovery* ± 1σ	QC Criteria (80-120%) **	Mean Recovery* ± 1σ	QC Criteria (80-120%) **
¹⁰⁷ Ag (328.068 nm)	2.5	6.0	5.3	11.5	106 ± 7%	Pass	107 ± 8%	Pass				
Al (396.152 nm)	12863.3	35208.9	102222.2	129044.4	NA	NA	NA	NA				
¹⁰⁹ Ag (188.980 nm)	62.5	7.3	114.7	298.8	104 ± 2%	Pass	116 ± 10%	Pass				
¹¹⁰ Ag (249.772 nm)	21288.9	4264.0	38900.0	18338.9	92 ± 235%	Pass	84 ± 46%	Pass				
¹¹¹ Ba (614.171 nm)	9706.9	5563.3	32433.3	31657.8	111 ± 57%	Pass	82 ± 26%	Pass				
¹¹² Be (313.042 nm)	1.1	2.5	34.0	19.7	100 ± 4%	Pass	109 ± 3%	Pass				
Ca (317.933 nm)	407444.4	209533.3	258333.3	724444.4	NA	NA	NA	NA				
¹¹⁴ Cd (214.439 nm)	451.5	43.1	31.5	37.6	100 ± 2%	Pass	105 ± 3%	Pass				
¹¹⁵ Ce	14.6	52.6	1104.9	722.7	115 ± 35%	Pass						
¹¹⁶ Co (238.892 nm)	586.7	41.8	786.5	44.5	105 ± 4%	Pass	95 ± 16%	Pass				
¹¹⁷ Cr (267.716 nm)	207.5	449.4	1639.6	3824.7	111 ± 14%	Pass	110 ± 9%	Pass				
¹¹⁸ Cs	6.0	199.3	36.5	209.0	103 ± 1%	Pass						
¹¹⁹ Cu (327.395 nm)	9460.7	9315.1	28900.0	14090.0	119 ± 55%	Pass	99 ± 22%	Pass				
¹²⁰ Dy	1.0	4.0	90.9	42.4	112 ± 23%	Pass						
¹²¹ Er	0.5	1.7	48.4	21.8	114 ± 24%	Pass						
¹²² Eu	0.3	1.9	27.3	8.6	116 ± 24%	Pass						
¹²³ Fe (238.204 nm)	17888.9	42313.3	110444.4	65013.3	NA	NA	NA	NA				
¹²⁴ Ga (294.363 nm)	4.6	13.6	277.5	324.9	107 ± 3%	Pass	110 ± 7%	Pass				
¹²⁵ Gd	1.2	5.1	107.4	47.6	112 ± 23%	Pass						
Ge (209.426 nm)	332.4	4335.4	0.0	4486.2	NA	NA	NA	NA				
¹²⁷ Hf	0.5	1.2	42.9	16.9	84 ± 17%	Pass						
¹²⁸ Hg	0.0	5.8	0.0	21.2	106 ± 2%	Pass						
¹²⁹ Ho	0.2	0.7	18.1	8.0	114 ± 23%	Pass						
¹³⁰ Ir (224.268 nm)	1083.6	299.4	382.0	213.8	101 ± 2%	Pass	108 ± 19%	Pass				
K (766.491 nm)	3657666.7	412522.2	3560111.1	4483111.1	NA	NA	NA	NA				
¹³² La	9.2	45.8	511.4	468.8	111 ± 27%	Pass						
¹³³ Li (670.783 nm)	3658.3	1027.7	902.3	1206.2	106 ± 4%	Pass	117 ± 17%	Pass				
¹³⁴ Ti	0.0	0.6	1.9	7.2	114 ± 24%	Pass						
Mg (280.270 nm)	264888.9	841077.8	7852000.0	1583555.6	NA	NA	NA	NA				
¹³⁵ Mn	13724.4	46072.2	89086.7	98117.8	NA	NA	NA	NA				
¹³⁶ Mo (202.032 nm)	309.8	396.3	1072.1	927.6	99 ± 2%	Pass	105 ± 4%	Pass				
Na (589.592 nm)	104466.7	15156.7	694644.4	149900.0	NA	NA	NA	NA				
¹³⁹ Nd	2.6	6.3	106.4	58.5	92 ± 9%	Pass						
¹⁴⁰ Ni	7.2	27.9	578.4	299.9	105 ± 25%	Pass						
¹⁴¹ Ni	843.7	374.6	1646.4	1259.2	108 ± 11%	Pass	110 ± 9%	Pass				
P (213.618 nm)	400000.0	175777.8	487922.2	111733.3	NA	NA	NA	NA				
¹⁴³ Pb (220.353 nm)	16.2	74.5	506.8	751.0	106 ± 3%	Pass	95 ± 21%	Pass				
¹⁴⁴ Pd	0.7	0.8	6.3	0.9	97 ± 3%	Pass	96 ± 9%	Pass				
¹⁴⁵ Pb	1.9	7.6	132.7	85.4	108 ± 23%	Pass						
¹⁴⁶ Rb	6738.0	34775.8	15829.5	20362.2	94 ± 35%	Pass						
¹⁴⁷ Rh	4.9	0.0	5.0	1.7	108 ± 3%	Pass						
¹⁴⁸ Sn	4.9	0.0	0.5	0.7	100 ± 3%	Pass						
S (180.669 nm)	2346777.8	842411.1	397622.2	1065888.9	NA	NA	NA	NA				
¹⁴⁹ Sb (220.353 nm)	2.9	2.6	17.4	55.7	95 ± 4%	Pass	102 ± 19%	Pass				
¹⁵⁰ Se	303.1	14.9	47.2	83.1	106 ± 4%	Pass						
¹⁵¹ Sm	54901.9	162327.7	348062.1	324615.4	NA	NA	NA	NA				
¹⁵² Sm	1.5	5.0	120.5	55.9	102 ± 22%	Pass						
¹⁵³ Sn (189.925 nm)	18.3	2099.0	57.6	189.0	98 ± 2%	Pass	104 ± 6%	Pass				
¹⁵⁴ Sm	36350.0	14511.1	214044.0	69097.8	108 ± 68%	Pass						
¹⁵⁵ Eu	1.8	19.1	93.8	509.7	106 ± 27%	Pass						
¹⁵⁶ Tb	393.7	1603.9	99519.9	12881.6	104 ± 3%	Pass						
¹⁵⁷ Tb	0.0	5.2	10.9	11.2	107 ± 2%	Pass						
¹⁵⁸ Tm	0.1	0.2	6.6	2.9	105 ± 24%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						
¹⁵⁹ Tm	22.1	3807.2	37.4	0.0	107 ± 3%	Pass						